
Midmark M11 - v1.0.0
Total Cycles: 3220

Name: RITTER 1
Sterilizer ID: V2178198
05/11/2023 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.7 F 0.0
2:00 154.9 F 0.0
4:00 182.5 F 1.4
6:00 201.1 F 6.0
8:00 229.4 F 10.8
10:00 226.0 F 5.7
12:00 237.6 F 11.1
14:00 233.9 F 8.0
16:00 243.2 F 11.5
18:00 245.7 F 12.9
20:00 258.0 F 19.4
22:00 269.0 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.7 F 28.8
2:00 271.5 F 28.0
3:00 272.7 F 29.0
4:00 272.0 F 28.4

Min 271.4 F 27.9
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.0 F 0.0
5:00 230.7 F 0.0
10:00 218.2 F 0.0
15:00 204.1 F 0.0
20:00 204.3 F 0.0
25:00 218.3 F 0.0
30:00 225.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 23:07
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:16
CYCLE COMPLETE

05/11/2023 01:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____