
Midmark M11 - v1.0.5
Total Cycles: 2492

Name: RITTER2
Sterilizer ID: V2313228
06/09/2023 08:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.0 F 0.0
2:00 146.8 F 0.0
4:00 149.4 F 1.5
6:00 177.0 F 5.3
8:00 205.6 F 10.1
10:00 225.3 F 7.1
12:00 230.7 F 10.1
14:00 234.0 F 8.5
16:00 242.0 F 12.5
18:00 240.3 F 11.0
20:00 253.0 F 17.3
22:00 264.0 F 23.8
24:00 271.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.4
1:00 271.7 F 29.3
2:00 271.9 F 29.3
3:00 272.2 F 29.4
4:00 272.4 F 29.6

Min 271.5 F 29.3
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.3 F 0.0
5:00 217.1 F 0.0
10:00 230.2 F 0.0
15:00 228.7 F 0.0
20:00 229.6 F 0.0
25:00 227.0 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 24:47
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:55
CYCLE COMPLETE

06/09/2023 09:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____