

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3402

Name: RITTER2  
Sterilizer ID: V2313228  
02/22/2024 11:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.5 F 0.0  
2:00 131.5 F 0.4  
4:00 141.4 F 2.5  
6:00 168.4 F 5.8  
8:00 191.7 F 9.9  
10:00 207.3 F 8.2  
12:00 218.1 F 7.6  
14:00 234.1 F 12.2  
16:00 233.9 F 9.4  
18:00 239.0 F 10.2  
20:00 246.3 F 14.0  
22:00 257.8 F 20.2  
24:00 268.6 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.3  
1:00 272.1 F 29.5  
2:00 271.8 F 29.9  
3:00 271.7 F 29.3  
4:00 271.8 F 29.9

Min 271.1 F 29.1  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.1 F 0.0  
5:00 214.3 F 0.0  
10:00 214.2 F 0.0  
15:00 193.2 F 0.0  
20:00 189.0 F 0.0  
25:00 185.8 F 0.0  
30:00 184.5 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 25:13  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:03:43  
CYCLE COMPLETE

02/22/2024 12:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_