
Midmark M11 - v1.0.0
Total Cycles: 1073

Name: RITTER 1
Sterilizer ID: V2178198
08/05/2021 08:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.3 F 0.0
2:00 123.9 F 0.0
4:00 150.6 F 1.9
6:00 174.9 F 4.8
8:00 194.8 F 9.1
10:00 218.7 F 8.3
12:00 227.8 F 7.4
14:00 238.0 F 12.1
16:00 234.6 F 8.4
18:00 243.6 F 11.8
20:00 244.2 F 12.1
22:00 256.1 F 18.4
24:00 267.2 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.7 F 28.7
2:00 271.4 F 28.0
3:00 272.6 F 28.8
4:00 271.9 F 28.1

Min 271.4 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.2 F 0.0
5:00 214.9 F 0.0
10:00 209.6 F 0.0
15:00 206.5 F 0.0
20:00 200.1 F 0.0
25:00 201.4 F 0.0
30:00 207.3 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 25:28
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:03:36
CYCLE COMPLETE

08/05/2021 09:29 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____