
Midmark M11 - v1.0.5
Total Cycles: 625

Name: RITTER2
Sterilizer ID: V2313228
12/13/2021 07:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 82.2 F 0.0
4:00 148.6 F 0.1
6:00 99.2 F 0.3
8:00 118.0 F 3.0
10:00 150.5 F 6.4
12:00 179.4 F 10.0
14:00 199.8 F 11.1
16:00 224.6 F 6.1
18:00 235.1 F 10.0
20:00 235.1 F 9.6
22:00 240.7 F 11.4
24:00 236.9 F 9.4
26:00 249.7 F 15.5
28:00 260.8 F 21.6
30:00 270.4 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.2
1:00 272.0 F 29.2
2:00 272.3 F 29.5
3:00 272.7 F 29.8
4:00 271.6 F 29.3

Min 271.5 F 29.1
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.5 F 0.0
5:00 237.3 F 0.0
10:00 225.8 F 0.0
15:00 228.9 F 0.0
20:00 222.2 F 0.0
25:00 219.6 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 31:33
STERILIZING: 4:00
VENTING: 3:00
DRYING: 30:00
TOTAL CYCLE: 01:09:40
CYCLE COMPLETE

12/13/2021 08:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____