
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001417
10:38 19/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:40
PRESSURIZING 3:41
STERILIZING 4:48
136.0 C 220kPa 4:48
136.4 C 220kPa 5:18
136.3 C 225kPa 5:48
136.4 C 224kPa 6:18
136.3 C 224kPa 6:48
136.4 C 225kPa 7:18
136.4 C 221kPa 7:48
136.1 C 222kPa 8:18
136.1 C 222kPa 8:18
VENTING 8:19
AIR DRYING 9:11
CYCLE COMPLETE 24:11

Digital Signature #
75249673864B2B4F
