

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2510

Name: RITTER2  
Sterilizer ID: V2313228  
06/13/2023 01:59 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.1 F 0.0  
2:00 197.8 F 4.1  
4:00 204.0 F 10.0  
6:00 222.0 F 6.4  
8:00 233.0 F 11.9  
10:00 235.3 F 9.9  
12:00 236.0 F 8.7  
14:00 250.9 F 16.5  
16:00 263.6 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 272.4 F 29.7  
2:00 272.0 F 29.9  
3:00 272.1 F 29.4  
4:00 272.3 F 30.1  
  
Min 271.5 F 29.2  
Max 272.6 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.7 F 0.0  
5:00 232.2 F 0.0  
10:00 224.1 F 0.0  
15:00 226.9 F 0.0  
20:00 228.6 F 0.0  
25:00 228.5 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 2:24  
HEATING: 17:57  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:57:38  
CYCLE COMPLETE

06/13/2023 02:56 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_