
Midmark M11 - v1.0.0
Total Cycles: 4467

Name: RITTER 1
Sterilizer ID: V2178198
04/17/2024 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 112.7 F 0.0
2:00 114.0 F 1.0
4:00 133.4 F 3.5
6:00 161.3 F 6.6
8:00 183.2 F 10.3
10:00 205.1 F 5.8
12:00 218.7 F 5.7
14:00 227.2 F 9.7
16:00 238.1 F 10.0
18:00 238.7 F 10.6
20:00 235.3 F 8.2
22:00 248.2 F 14.1
24:00 259.2 F 20.2
26:00 269.4 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.0 F 28.6
2:00 272.8 F 28.9
3:00 271.8 F 28.9
4:00 271.8 F 28.5

Min 271.4 F 28.4
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.1 F 0.0
5:00 226.4 F 0.0
10:00 204.2 F 0.0
15:00 214.0 F 0.0
20:00 196.0 F 0.0
25:00 192.6 F 0.0
30:00 218.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 27:42
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:06:16
CYCLE COMPLETE

04/17/2024 02:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____