
Midmark M11 - v1.0.0
Total Cycles: 3731

Name: RITTER 1
Sterilizer ID: V2178198
09/07/2023 09:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.6 F 0.0
2:00 175.5 F 0.0
4:00 173.7 F 1.3
6:00 190.1 F 4.1
8:00 207.2 F 7.8
10:00 216.4 F 12.1
12:00 226.3 F 6.8
14:00 237.9 F 11.6
16:00 235.1 F 8.8
18:00 241.0 F 10.9
20:00 245.9 F 13.2
22:00 257.3 F 19.3
24:00 267.6 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.4 F 29.0
2:00 272.0 F 28.5
3:00 271.1 F 28.3
4:00 272.2 F 29.1

Min 271.0 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 219.1 F 0.0
10:00 226.3 F 0.0
15:00 225.6 F 0.0
20:00 230.0 F 0.0
25:00 228.4 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 25:30
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:04:09
CYCLE COMPLETE

09/07/2023 10:22 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____