
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001279
9:48 30/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:10
PRESSURIZING 4:11
STERILIZING 5:29
136.0 C 222kPa 5:29
136.4 C 224kPa 5:59
136.3 C 223kPa 6:29
136.4 C 224kPa 6:59
136.2 C 222kPa 7:29
136.4 C 224kPa 7:59
136.4 C 224kPa 8:29
136.1 C 222kPa 8:59
136.1 C 222kPa 8:59
VENTING 8:59
AIR DRYING 9:56
CYCLE COMPLETE 24:56

Digital Signature #
FB9F4D6FB891CF57
