
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.8uS / 0.5ppm
CYCLE NUMBER 000634
10:17 08/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:56
STERILIZING 4:06
136.0 C 226kPa 4:06
136.4 C 232kPa 4:36
136.3 C 230kPa 5:06
136.1 C 229kPa 5:36
136.4 C 231kPa 6:06
136.4 C 231kPa 6:36
136.3 C 230kPa 7:06
136.4 C 232kPa 7:36
136.4 C 232kPa 7:36
VENTING 7:36
AIR DRYING 8:27
CYCLE COMPLETE 23:27

Digital Signature #
BB3B1142070D484F
