

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 005754  
10:00 14/06/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:45  
PRESSURIZING 2:52  
STERILIZING 3:57  
136.0 C 220kPa 3:57  
136.4 C 224kPa 4:27  
136.1 C 222kPa 4:57  
136.4 C 221kPa 5:27  
136.3 C 222kPa 5:57  
136.2 C 223kPa 6:27  
136.2 C 223kPa 6:57  
136.2 C 221kPa 7:27  
136.2 C 221kPa 7:27  
VENTING 7:27  
AIR DRYING 8:14  
CYCLE COMPLETE 8:31  
-----

Digital Signature #  
6416CC4C98123499  
-----