
Midmark M11 - v1.0.0
Total Cycles: 2534

Name: RITTER 1
Sterilizer ID: V2178198
11/09/2022 08:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 79.0 F 0.0
2:00 79.8 F 0.0
4:00 99.4 F 0.0
6:00 135.1 F 0.3
8:00 155.3 F 3.0
10:00 174.8 F 6.7
12:00 197.7 F 10.6
14:00 221.9 F 5.1
16:00 225.2 F 6.7
18:00 233.6 F 10.7
20:00 232.2 F 7.5
22:00 241.6 F 11.9
24:00 239.0 F 9.8
26:00 252.0 F 16.2
28:00 263.0 F 22.6
30:00 272.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 28.8
2:00 271.5 F 28.2
3:00 272.5 F 28.8
4:00 271.7 F 28.1

Min 271.5 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 204.4 F 0.0
10:00 199.2 F 0.0
15:00 192.4 F 0.0
20:00 190.7 F 0.0
25:00 195.5 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:17
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:15
CYCLE COMPLETE

11/09/2022 09:40 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____