
Midmark M11 - v1.0.0
Total Cycles: 2117

Name: RITTER 1
Sterilizer ID: V2178198
07/12/2022 02:24 PM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.5 F 0.0
2:00 132.9 F 0.5
4:00 150.4 F 2.7
6:00 179.1 F 6.7
8:00 202.6 F 12.0
10:00 220.3 F 6.2
12:00 234.7 F 11.7
14:00 235.0 F 8.8
16:00 238.7 F 9.5
18:00 247.4 F 13.9
20:00 260.2 F 20.9
22:00 271.6 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 271.5 F 28.1
2:00 271.7 F 28.1
3:00 272.1 F 28.4

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 226.1 F 0.0
Dry Cycle Aborted

FILLING: 0:60
HEATING: 22:30
STERILIZING: 3:00
VENTING: 2:47
DRYING: 0:17
TOTAL CYCLE: 00:29:34

07/12/2022 02:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____