
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006767
14:04 19/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:12
STERILIZING 4:16
136.4 C 221kPa 4:16
136.3 C 224kPa 4:46
136.2 C 221kPa 5:16
136.4 C 224kPa 5:46
136.1 C 221kPa 6:16
136.2 C 223kPa 6:46
136.3 C 225kPa 7:16
136.2 C 219kPa 7:46
136.2 C 219kPa 7:46
VENTING 7:46
AIR DRYING 8:44
CYCLE COMPLETE 8:50

Digital Signature #
575B1D267B3EA94F
