
Midmark M11 - v1.0.0
Total Cycles: 3664

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2023 02:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.9 F 0.0
2:00 113.7 F 1.1
4:00 137.0 F 3.7
6:00 168.5 F 7.0
8:00 185.6 F 10.7
10:00 216.2 F 4.0
12:00 220.9 F 7.1
14:00 231.3 F 11.5
16:00 231.1 F 7.4
18:00 242.1 F 12.8
20:00 241.0 F 10.9
22:00 252.9 F 17.0
24:00 263.7 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 271.5 F 28.4
2:00 271.5 F 28.5
3:00 271.8 F 28.7
4:00 272.1 F 28.9

Min 271.4 F 28.4
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.8 F 0.0
5:00 235.5 F 0.0
10:00 211.7 F 0.0
15:00 211.4 F 0.0
20:00 204.4 F 0.0
25:00 199.8 F 0.0
30:00 187.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:59
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:04:21
CYCLE COMPLETE

08/16/2023 03:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____