
Midmark M11 - v1.0.5
Total Cycles: 2208

Name: RITTER2
Sterilizer ID: V2313228
04/03/2023 01:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.1 F 0.0
2:00 182.6 F 4.0
4:00 199.5 F 8.8
6:00 220.5 F 8.0
8:00 231.6 F 10.3
10:00 231.9 F 7.9
12:00 239.8 F 10.7
14:00 247.3 F 14.4
16:00 260.2 F 21.5
18:00 271.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.1
1:00 272.4 F 29.6
2:00 271.3 F 29.1
3:00 272.3 F 29.7
4:00 271.7 F 29.1

Min 271.3 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 237.2 F 0.0
10:00 221.3 F 0.0
15:00 218.2 F 0.0
20:00 206.1 F 0.0
25:00 201.4 F 0.0
30:00 200.6 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 18:35
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:49
CYCLE COMPLETE

04/03/2023 02:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____