
Midmark M11 - v1.0.0
Total Cycles: 3409

Name: RITTER 1
Sterilizer ID: V2178198
06/19/2023 04:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.0 F 0.0
2:00 176.1 F 1.5
4:00 182.4 F 5.1
6:00 205.1 F 9.4
8:00 226.0 F 8.4
10:00 233.7 F 9.2
12:00 235.6 F 10.3
14:00 243.0 F 11.9
16:00 240.5 F 10.4
18:00 253.0 F 16.6
20:00 263.8 F 22.9
22:00 272.2 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.0
1:00 272.9 F 29.2
2:00 272.1 F 29.0
3:00 272.1 F 28.6
4:00 272.6 F 28.9

Min 271.0 F 27.9
Max 273.1 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 219.0 F 0.0
10:00 231.0 F 0.0
15:00 227.8 F 0.0
20:00 227.1 F 0.0
25:00 227.2 F 0.0
30:00 226.8 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 22:13
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:01:09
CYCLE COMPLETE

06/19/2023 05:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____