
Midmark M11 - v1.0.5
Total Cycles: 1236

Name: RITTER2
Sterilizer ID: V2313228
06/28/2022 01:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.4 F 0.0
2:00 141.7 F 1.7
4:00 151.1 F 5.1
6:00 176.3 F 9.3
8:00 204.8 F 7.1
10:00 223.3 F 8.5
12:00 241.3 F 13.5
14:00 236.7 F 10.5
16:00 237.9 F 8.4
18:00 247.7 F 15.1
20:00 259.2 F 21.5
22:00 270.3 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.6
1:00 272.4 F 30.5
2:00 273.0 F 30.3
3:00 272.3 F 29.9
4:00 271.8 F 29.5

Min 271.6 F 29.5
Max 273.9 F 30.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 222.0 F 0.0
10:00 225.3 F 0.0
15:00 218.6 F 0.0
20:00 215.0 F 0.0
25:00 211.8 F 0.0
30:00 207.9 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 23:02
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:43
CYCLE COMPLETE

06/28/2022 02:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____