
Midmark M11 - v1.0.0
Total Cycles: 1130

Name: RITTER 1
Sterilizer ID: V2178198
08/30/2021 11:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.1 F 0.0
2:00 181.7 F 3.2
4:00 204.8 F 8.6
6:00 220.7 F 8.3
8:00 234.6 F 9.6
10:00 235.6 F 9.1
12:00 245.7 F 13.4
14:00 244.7 F 12.6
16:00 258.4 F 19.8
18:00 270.0 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 28.8
2:00 271.7 F 28.5
3:00 272.3 F 28.7
4:00 271.5 F 28.4

Min 271.3 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.4 F 0.0
5:00 228.8 F 0.0
10:00 205.1 F 0.0
15:00 205.4 F 0.0
20:00 203.8 F 0.0
25:00 214.0 F 0.0
30:00 229.8 F 0.0

Drying Complete

FILLING: 1:53
HEATING: 18:52
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:52
CYCLE COMPLETE

08/30/2021 12:39 PM
ACCEPTED: X
REJECTED:
Operator: U02