

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1440

Name: RITTER2  
Sterilizer ID: V2313228  
08/31/2022 01:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.0 F 0.0  
2:00 182.7 F 2.5  
4:00 198.0 F 7.5  
6:00 216.1 F 13.1  
8:00 229.6 F 9.3  
10:00 234.8 F 9.5  
12:00 244.4 F 13.0  
14:00 244.8 F 13.0  
16:00 257.3 F 19.9  
18:00 268.7 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.9  
1:00 272.4 F 29.6  
2:00 271.5 F 28.9  
3:00 271.8 F 29.5  
4:00 272.2 F 29.6  
  
Min 271.1 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.4 F 0.0  
5:00 219.9 F 0.0  
10:00 239.6 F 0.0  
15:00 234.6 F 0.0  
20:00 225.7 F 0.0  
25:00 223.1 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 2:01  
HEATING: 19:10  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:19  
CYCLE COMPLETE

08/31/2022 02:01 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_