
Midmark M11 - v1.0.5
Total Cycles: 4495

Name: RITTER2
Sterilizer ID: V2313228
04/08/2025 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 70.4 F 0.0
4:00 71.5 F 0.3
6:00 81.8 F 2.0
8:00 139.8 F 4.4
10:00 148.5 F 7.1
12:00 167.2 F 10.2
14:00 184.8 F 13.4
16:00 213.1 F 4.7
18:00 217.9 F 7.8
20:00 227.2 F 11.4
22:00 229.9 F 6.9
24:00 239.7 F 11.8
26:00 236.9 F 9.6
28:00 248.3 F 15.0
30:00 258.2 F 20.5
32:00 267.3 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.0 F 30.2
2:00 271.9 F 29.8
3:00 272.5 F 30.1
4:00 272.2 F 30.4

Min 270.8 F 29.1
Max 273.1 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 216.7 F 0.0
10:00 207.9 F 0.0
15:00 181.2 F 0.0
20:00 173.3 F 0.0
25:00 173.8 F 0.0
30:00 173.9 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 33:34
STERILIZING: 4:00
VENTING: 4:01
DRYING: 30:00
TOTAL CYCLE: 01:12:25
CYCLE COMPLETE

04/08/2025 11:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____