
Midmark M11 - v1.0.5
Total Cycles: 615

Name: RITTER2
Sterilizer ID: V2313228
12/08/2021 03:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.4 F 0.0
2:00 154.4 F 3.1
4:00 181.9 F 6.8
6:00 199.6 F 11.6
8:00 218.0 F 5.6
10:00 231.6 F 10.5
12:00 231.5 F 8.0
14:00 243.6 F 13.4
16:00 243.4 F 12.5
18:00 255.8 F 18.9
20:00 266.8 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.4 F 29.6
2:00 271.4 F 28.8
3:00 271.8 F 29.6
4:00 271.8 F 29.4

Min 271.0 F 28.7
Max 272.5 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.1 F 0.0
5:00 207.8 F 0.0
10:00 197.6 F 0.0
15:00 198.6 F 0.0
20:00 197.0 F 0.0
25:00 195.9 F 0.0
30:00 194.3 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:06
CYCLE COMPLETE

12/08/2021 04:51 PM
ACCEPTED: X
REJECTED:
Operator: U02