
Midmark M11 - v1.0.5
Total Cycles: 2431

Name: RITTER2
Sterilizer ID: V2313228
05/29/2023 03:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.7 F 0.0
2:00 191.4 F 3.8
4:00 199.0 F 8.6
6:00 215.5 F 9.1
8:00 227.1 F 9.5
10:00 235.2 F 8.8
12:00 242.5 F 13.0
14:00 242.8 F 12.0
16:00 255.7 F 18.8
18:00 267.1 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.5 F 29.8
2:00 271.9 F 29.1
3:00 271.5 F 29.3
4:00 272.5 F 29.7

Min 271.2 F 28.8
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.8 F 0.0
5:00 217.8 F 0.0
10:00 229.2 F 0.0
15:00 230.3 F 0.0
20:00 227.0 F 0.0
25:00 226.2 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:27
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:32
CYCLE COMPLETE

05/29/2023 04:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____