

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 001793  
7:50 17/06/2021

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:18  
PRESSURIZING 3:19  
STERILIZING 4:20  
136.4 C 221kPa 4:20  
136.3 C 223kPa 4:50  
136.4 C 219kPa 5:20  
136.3 C 222kPa 5:50  
136.1 C 222kPa 6:20  
136.4 C 223kPa 6:50  
136.4 C 221kPa 7:20  
136.1 C 222kPa 7:50  
136.1 C 222kPa 7:50  
VENTING 7:50  
AIR DRYING 8:40  
CYCLE COMPLETE 8:49  
-----

Digital Signature #  
1619E9AB06831ACE  
-----