

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4194

Name: RITTER2  
Sterilizer ID: V2313228  
12/06/2024 01:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.7 F 0.0  
2:00 129.5 F 0.1  
4:00 140.2 F 0.8  
6:00 169.0 F 3.5  
8:00 193.7 F 6.8  
10:00 214.2 F 10.7  
12:00 224.7 F 6.5  
14:00 232.2 F 8.6  
16:00 241.6 F 13.1  
18:00 239.9 F 10.8  
20:00 237.1 F 9.3  
22:00 250.0 F 15.5  
24:00 260.8 F 21.6  
26:00 270.6 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.1  
1:00 272.9 F 29.3  
2:00 272.4 F 29.7  
3:00 271.8 F 29.0  
4:00 272.5 F 29.7

Min 271.6 F 29.0  
Max 273.0 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.4 F 0.0  
5:00 241.8 F 0.0  
10:00 181.1 F 0.0  
15:00 174.8 F 0.0  
20:00 169.7 F 0.0  
25:00 166.9 F 0.0  
30:00 165.1 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 27:38  
STERILIZING: 4:00  
VENTING: 3:59  
DRYING: 30:00  
TOTAL CYCLE: 01:06:35  
CYCLE COMPLETE

12/06/2024 02:41 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_