

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000445  
7:12 01/02/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:18  
PRESSURIZING 3:19  
STERILIZING 4:29  
136.0 C 228kPa 4:29  
136.0 C 230kPa 4:59  
136.2 C 231kPa 5:29  
136.3 C 232kPa 5:59  
136.4 C 233kPa 6:29  
136.3 C 229kPa 6:59  
136.4 C 232kPa 7:29  
136.3 C 231kPa 7:59  
136.3 C 231kPa 7:59  
VENTING 7:59  
AIR DRYING 8:53  
CYCLE COMPLETE 11:22

-----  
Digital Signature #  
28A0CB28222ABC55  
-----