
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001245
14:32 02/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:59
PRESSURIZING 4:00
STERILIZING 5:10
136.4 C 223kPa 5:10
136.3 C 223kPa 5:40
136.4 C 225kPa 6:10
136.3 C 222kPa 6:40
136.4 C 224kPa 7:10
136.4 C 221kPa 7:40
136.4 C 224kPa 8:10
136.4 C 224kPa 8:40
136.4 C 224kPa 8:40
VENTING 8:40
AIR DRYING 9:33
CYCLE COMPLETE 24:33

Digital Signature #
4EDE03412028C864
