
Midmark M11 - v1.0.0
Total Cycles: 4690

Name: RITTER 1
Sterilizer ID: V2178198
06/28/2024 02:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.1 F 0.0
2:00 129.8 F 1.9
4:00 145.7 F 4.9
6:00 169.7 F 8.5
8:00 196.4 F 13.0
10:00 213.9 F 6.6
12:00 228.5 F 11.6
14:00 229.4 F 7.9
16:00 243.1 F 12.6
18:00 242.1 F 11.9
20:00 254.8 F 18.4
22:00 266.1 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.3
2:00 271.7 F 29.2
3:00 272.4 F 29.2
4:00 271.6 F 29.1

Min 271.4 F 28.8
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.2 F 0.0
5:00 215.8 F 0.0
10:00 220.9 F 0.0
15:00 220.3 F 0.0
20:00 222.8 F 0.0
25:00 224.1 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 23:37
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:02:09
CYCLE COMPLETE

06/28/2024 03:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____