
Midmark M11 - v1.0.5
Total Cycles: 2416

Name: RITTER2
Sterilizer ID: V2313228
05/25/2023 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.1 F 0.0
2:00 188.8 F 3.8
4:00 195.7 F 8.8
6:00 225.3 F 7.9
8:00 228.6 F 10.0
10:00 232.8 F 8.0
12:00 242.7 F 11.8
14:00 246.0 F 13.5
16:00 258.9 F 20.7
18:00 270.6 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.5 F 29.5
2:00 271.5 F 29.1
3:00 272.7 F 29.6
4:00 271.5 F 28.9

Min 271.4 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.8 F 0.0
5:00 218.2 F 0.0
10:00 244.9 F 0.0
15:00 230.9 F 0.0
20:00 226.9 F 0.0
25:00 228.4 F 0.0
30:00 230.5 F 0.0

Drying Complete

FILLING: 1:55
HEATING: 18:49
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:57:55
CYCLE COMPLETE

05/25/2023 01:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____