
Midmark M11 - v1.0.0
Total Cycles: 3573

Name: RITTER 1
Sterilizer ID: V2178198
07/24/2023 06:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.8 F 0.0
2:00 153.2 F 0.5
4:00 171.1 F 2.8
6:00 193.7 F 6.0
8:00 208.6 F 9.8
10:00 220.6 F 9.8
12:00 227.4 F 8.2
14:00 238.0 F 12.7
16:00 237.5 F 10.0
18:00 237.9 F 9.2
20:00 248.6 F 14.4
22:00 259.7 F 20.6
24:00 270.0 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.4 F 28.7
2:00 271.5 F 28.4
3:00 272.3 F 28.8
4:00 271.3 F 28.2

Min 271.3 F 28.1
Max 272.5 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.3 F 0.0
5:00 201.1 F 0.0
10:00 207.4 F 0.0
15:00 208.5 F 0.0
20:00 204.2 F 0.0
25:00 204.7 F 0.0
30:00 206.8 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:55
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:03:37
CYCLE COMPLETE

07/24/2023 07:12 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____