
Midmark M11 - v1.0.5
Total Cycles: 1474

Name: RITTER2
Sterilizer ID: V2313228
09/12/2022 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.2 F 0.0
2:00 176.6 F 2.9
4:00 193.9 F 6.3
6:00 206.3 F 10.6
8:00 216.0 F 5.3
10:00 229.4 F 10.5
12:00 231.2 F 7.8
14:00 242.6 F 13.4
16:00 242.3 F 12.1
18:00 254.6 F 18.6
20:00 265.9 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.3
1:00 272.5 F 29.9
2:00 271.5 F 29.1
3:00 272.2 F 30.1
4:00 272.2 F 29.6

Min 271.2 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 225.2 F 0.0
10:00 231.2 F 0.0
15:00 226.5 F 0.0
20:00 225.1 F 0.0
25:00 224.0 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 21:42
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

09/12/2022 11:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____