
Midmark M11 - v1.0.5
Total Cycles: 3539

Name: RITTER2
Sterilizer ID: V2313228
04/17/2024 05:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.2 F 0.0
2:00 176.7 F 3.5
4:00 195.1 F 8.9
6:00 221.8 F 8.0
8:00 228.4 F 11.4
10:00 234.8 F 9.3
12:00 236.7 F 9.0
14:00 249.4 F 15.6
16:00 262.0 F 23.1
18:00 271.6 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.2
1:00 272.3 F 29.7
2:00 271.1 F 29.2
3:00 272.3 F 29.8
4:00 271.4 F 29.1

Min 271.0 F 29.0
Max 272.8 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.3 F 0.0
5:00 221.7 F 0.0
10:00 196.0 F 0.0
15:00 191.6 F 0.0
20:00 187.7 F 0.0
25:00 185.3 F 0.0
30:00 184.0 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 18:08
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:23
CYCLE COMPLETE

04/17/2024 06:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____