
Midmark M11 - v1.0.5
Total Cycles: 1045

Name: RITTER2
Sterilizer ID: V2313228
04/28/2022 08:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.3 F 0.0
2:00 166.6 F 2.4
4:00 169.4 F 5.8
6:00 189.8 F 10.4
8:00 215.4 F 5.0
10:00 225.3 F 9.5
12:00 234.1 F 9.3
14:00 240.9 F 12.4
16:00 240.6 F 11.1
18:00 253.8 F 17.8
20:00 265.2 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.4 F 29.5
2:00 271.9 F 29.7
3:00 272.0 F 29.3
4:00 271.9 F 29.7

Min 271.5 F 29.0
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.5 F 0.0
5:00 213.4 F 0.0
10:00 224.6 F 0.0
15:00 226.8 F 0.0
20:00 220.6 F 0.0
25:00 217.5 F 0.0
30:00 214.9 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:05
CYCLE COMPLETE

04/28/2022 09:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____