
Midmark M11 - v1.0.0
Total Cycles: 4574

Name: RITTER 1
Sterilizer ID: V2178198
05/23/2024 07:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.7 F 0.0
2:00 73.4 F 0.2
4:00 109.9 F 1.2
6:00 120.7 F 3.4
8:00 145.5 F 6.1
10:00 169.3 F 9.3
12:00 187.2 F 12.7
14:00 202.1 F 3.0
16:00 213.2 F 6.2
18:00 223.3 F 9.5
20:00 233.5 F 13.4
22:00 234.0 F 9.0
24:00 241.7 F 11.5
26:00 242.8 F 11.9
28:00 254.2 F 17.7
30:00 264.4 F 23.7
32:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.8 F 29.2
2:00 271.8 F 28.4
3:00 272.6 F 29.3
4:00 272.3 F 28.9

Min 271.6 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 207.4 F 0.0
10:00 209.8 F 0.0
15:00 219.3 F 0.0
20:00 220.0 F 0.0
25:00 218.5 F 0.0
30:00 220.6 F 0.0

Drying Complete

FILLING: 0:48
HEATING: 32:10
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:10:20
CYCLE COMPLETE

05/23/2024 09:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____