

-----  
Midmark M11 - v1.0.5  
Total Cycles: 455

Name: RITTER2  
Sterilizer ID: V2313228  
10/18/2021 07:44 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.9 F 0.0  
2:00 74.0 F 0.0  
4:00 99.2 F 0.0  
6:00 113.7 F 0.1  
8:00 149.8 F 0.2  
10:00 167.2 F 2.3  
12:00 183.9 F 5.5  
14:00 199.4 F 8.9  
16:00 212.9 F 12.5  
18:00 221.6 F 6.8  
20:00 232.5 F 11.0  
22:00 232.8 F 8.2  
24:00 243.1 F 13.3  
26:00 242.4 F 12.1  
28:00 254.2 F 18.2  
30:00 264.6 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.2  
1:00 271.9 F 29.3  
2:00 272.3 F 29.7  
3:00 271.3 F 29.5  
4:00 271.7 F 29.3

Min 271.1 F 29.0  
Max 272.3 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.1 F 0.0  
5:00 221.2 F 0.0  
10:00 226.4 F 0.0  
15:00 228.8 F 0.0  
20:00 203.4 F 0.0  
25:00 198.3 F 0.0  
30:00 227.5 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 31:56  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 01:10:04  
CYCLE COMPLETE

10/18/2021 08:54 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01