
Midmark M11 - v1.0.0
Total Cycles: 4910

Name: RITTER 1
Sterilizer ID: V2178198
09/16/2024 08:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.2 F 0.0
2:00 71.1 F 0.0
4:00 101.7 F 0.0
6:00 131.4 F 0.0
8:00 155.6 F 0.2
10:00 178.9 F 0.4
12:00 199.7 F 1.0
14:00 209.6 F 4.3
16:00 218.6 F 7.1
18:00 227.5 F 10.4
20:00 232.0 F 9.1
22:00 236.9 F 9.9
24:00 238.1 F 9.1
26:00 246.4 F 13.2
28:00 244.0 F 11.8
30:00 256.6 F 18.3
32:00 267.7 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.8
1:00 272.8 F 28.5
2:00 271.6 F 27.6
3:00 272.5 F 28.5
4:00 272.4 F 28.1

Min 271.5 F 27.6
Max 272.8 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 190.1 F 0.0
10:00 202.5 F 0.0
15:00 209.3 F 0.0
20:00 215.4 F 0.0
25:00 215.2 F 0.0
30:00 214.4 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 33:22
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:11:28
CYCLE COMPLETE

09/16/2024 09:44 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____