
Midmark M11 - v1.0.5
Total Cycles: 1688

Name: RITTER2
Sterilizer ID: V2313228
11/07/2022 12:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.7 F 0.0
2:00 161.3 F 0.2
4:00 191.8 F 2.1
6:00 205.6 F 7.2
8:00 224.9 F 12.7
10:00 231.2 F 9.2
12:00 236.6 F 9.6
14:00 243.3 F 12.0
16:00 246.5 F 13.8
18:00 259.5 F 21.0
20:00 271.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.9
1:00 272.7 F 29.6
2:00 271.5 F 28.9
3:00 272.5 F 29.4
4:00 271.5 F 28.9

Min 271.4 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.8 F 0.0
5:00 222.3 F 0.0
10:00 236.6 F 0.0
15:00 228.1 F 0.0
20:00 230.6 F 0.0
25:00 219.7 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 20:30
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:39
CYCLE COMPLETE

11/07/2022 01:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____