
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001917
9:01 07/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:03
STERILIZING 3:04
136.0 C 220kPa 3:04
136.4 C 225kPa 3:34
136.4 C 226kPa 4:04
136.2 C 223kPa 4:34
136.4 C 226kPa 5:04
136.3 C 224kPa 5:34
136.4 C 225kPa 6:04
136.1 C 223kPa 6:34
136.1 C 223kPa 6:34
VENTING 6:34
AIR DRYING 7:25
CYCLE COMPLETE 37:25

Digital Signature #
E6B00E770E468AEE
