
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001520
15:57 19/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:48
STERILIZING 3:55
136.5 C 225kPa 3:55
136.1 C 224kPa 4:25
136.1 C 223kPa 4:55
136.0 C 222kPa 5:25
136.4 C 225kPa 5:55
136.4 C 225kPa 6:25
136.4 C 225kPa 6:55
136.4 C 225kPa 7:25
136.4 C 225kPa 7:25
VENTING 7:25
AIR DRYING 8:18
CYCLE COMPLETE 8:24

Digital Signature #
4ADA3EDA4927ADF5
