

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5135

Name: RITTER 1  
Sterilizer ID: V2178198  
11/27/2024 11:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	159.2 F	0.0
2:00	157.0 F	0.1
4:00	163.1 F	0.3
6:00	193.0 F	1.0
8:00	209.0 F	4.7
10:00	223.3 F	8.7
12:00	235.5 F	13.3
14:00	235.4 F	9.8
16:00	237.0 F	8.9
18:00	245.1 F	12.4
20:00	245.7 F	13.0
22:00	258.8 F	20.0
24:00	269.8 F	27.0

STERILIZING

mm:ss	Degrees	PSI
0:00	271.3 F	28.2
1:00	272.4 F	28.9
2:00	272.1 F	28.5
3:00	271.3 F	28.2
4:00	272.2 F	28.9

Min 271.3 F 28.0  
Max 272.5 F 29.0

VENTING CHAMBER

mm:ss	Degrees	PSI
0:00	197.8 F	0.0
5:00	198.0 F	0.0
10:00	199.7 F	0.0
15:00	209.7 F	0.0
20:00	219.8 F	0.0
25:00	218.4 F	0.0
30:00	231.3 F	0.0

Drying Complete

FILLING: 0:47  
HEATING: 25:02  
STERILIZING: 4:00  
VENTING: 3:45  
DRYING: 30:00  
TOTAL CYCLE: 01:03:34  
CYCLE COMPLETE

11/27/2024 12:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_