
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002491
14:27 07/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:21
STERILIZING 3:21
136.0 C 220kPa 3:21
135.9 C 222kPa 3:51
136.3 C 225kPa 4:21
136.4 C 226kPa 4:51
136.4 C 225kPa 5:21
136.4 C 225kPa 5:51
136.1 C 222kPa 6:21
136.2 C 222kPa 6:51
136.2 C 222kPa 6:51
VENTING 6:51
AIR DRYING 7:41
CYCLE COMPLETE 26:58

Digital Signature #
F9BF5DD6C6129287
