
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 002264
16:14 14/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:07
STERILIZING 3:11
136.4 C 220kPa 3:11
136.4 C 225kPa 3:41
136.4 C 222kPa 4:11
136.3 C 224kPa 4:41
136.4 C 224kPa 5:11
136.2 C 223kPa 5:41
136.4 C 224kPa 6:11
136.4 C 224kPa 6:41
136.4 C 224kPa 6:41
VENTING 6:41
AIR DRYING 7:33
CYCLE COMPLETE 37:33

Digital Signature #
A20D6B3D7C487930
