

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4096

Name: RITTER 1  
Sterilizer ID: V2178198  
12/08/2023 09:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 68.9 F 0.0  
2:00 69.3 F 0.2  
4:00 72.4 F 1.0  
6:00 95.2 F 2.8  
8:00 130.4 F 5.1  
10:00 155.5 F 7.9  
12:00 171.1 F 11.0  
14:00 212.6 F 5.4  
16:00 207.6 F 4.5  
18:00 214.8 F 7.6  
20:00 224.9 F 10.8  
22:00 231.0 F 7.8  
24:00 236.7 F 10.3  
26:00 236.7 F 9.1  
28:00 245.1 F 12.9  
30:00 255.5 F 18.3  
32:00 265.0 F 24.0  
34:00 271.6 F 28.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.5 F 29.2  
2:00 271.8 F 28.5  
3:00 271.4 F 28.7  
4:00 272.6 F 29.1  
  
Min 271.1 F 28.3  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.3 F 0.0  
5:00 203.0 F 0.0  
10:00 196.9 F 0.0  
15:00 198.7 F 0.0  
20:00 210.4 F 0.0  
25:00 213.1 F 0.0  
30:00 210.2 F 0.0  
Drying Complete

FILLING: 0:36  
HEATING: 34:07  
STERILIZING: 4:00  
VENTING: 3:43  
DRYING: 30:00  
TOTAL CYCLE: 01:12:26  
CYCLE COMPLETE

12/08/2023 10:45 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_