

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.6uS / 0.3ppm  
CYCLE NUMBER 001850  
13:14 15/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:05  
PRESSURIZING 3:07  
STERILIZING 4:16  
136.0 C 221kPa 4:16  
136.3 C 223kPa 4:46  
136.3 C 225kPa 5:16  
136.4 C 223kPa 5:46  
136.4 C 223kPa 6:16  
136.4 C 224kPa 6:46  
136.3 C 223kPa 7:16  
136.4 C 226kPa 7:46  
136.4 C 226kPa 7:46  
VENTING 7:47  
AIR DRYING 8:40  
CYCLE COMPLETE 8:45

-----  
Digital Signature #  
FCEC2989F5DB4139  
-----