
Midmark M11 - v1.0.0
Total Cycles: 4051

Name: RITTER 1
Sterilizer ID: V2178198
11/22/2023 02:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 92.7 F 0.0
2:00 93.4 F 0.6
4:00 104.9 F 2.2
6:00 137.8 F 4.9
8:00 164.9 F 8.4
10:00 221.2 F 12.7
12:00 220.3 F 4.4
14:00 227.4 F 8.1
16:00 236.1 F 12.1
18:00 232.8 F 7.9
20:00 244.5 F 13.2
22:00 241.2 F 11.0
24:00 252.9 F 16.9
26:00 263.6 F 23.1
28:00 271.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 271.7 F 28.6
2:00 272.5 F 28.9
3:00 272.7 F 29.2
4:00 271.5 F 29.0

Min 271.2 F 28.6
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.0 F 0.0
5:00 231.9 F 0.0
10:00 223.6 F 0.0
15:00 223.1 F 0.0
20:00 226.8 F 0.0
25:00 216.1 F 0.0
30:00 198.9 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 28:52
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:07:21
CYCLE COMPLETE

11/22/2023 03:11 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____