
Midmark M11 - v1.0.0
Total Cycles: 2631

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2022 03:29 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.1 F 0.0
2:00 189.4 F 4.8
4:00 201.1 F 9.9
6:00 215.1 F 6.0
8:00 228.4 F 10.3
10:00 230.6 F 7.6
12:00 244.4 F 13.2
14:00 243.8 F 12.4
16:00 256.3 F 19.0
18:00 267.8 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.3 F 29.3
2:00 271.5 F 28.6
3:00 271.8 F 29.2
4:00 272.1 F 29.2

Min 271.1 F 28.4
Max 272.4 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 238.7 F 0.0
10:00 229.4 F 0.0
15:00 231.2 F 0.0
20:00 228.3 F 0.0
25:00 229.0 F 0.0
30:00 229.1 F 0.0

Drying Complete

FILLING: 1:33
HEATING: 19:17
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

11/30/2022 04:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____