
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001516
9:13 19/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:46
STERILIZING 3:51
136.0 C 220kPa 3:51
136.0 C 221kPa 4:21
136.2 C 221kPa 4:51
136.4 C 222kPa 5:21
136.0 C 221kPa 5:51
136.4 C 225kPa 6:21
136.4 C 225kPa 6:51
136.0 C 221kPa 7:21
136.0 C 221kPa 7:21
VENTING 7:22
AIR DRYING 8:11
CYCLE COMPLETE 9:35

Digital Signature #
81823856FBE5FDA0
