

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004300  
11:24 06/09/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:20  
PRESSURIZING 1:23  
STERILIZING 2:26  
136.0 C 220kPa 2:26  
136.4 C 223kPa 2:56  
136.2 C 220kPa 3:26  
136.4 C 224kPa 3:56  
136.4 C 224kPa 4:26  
136.2 C 221kPa 4:56  
135.8 C 218kPa 5:26  
136.4 C 221kPa 5:56  
136.4 C 221kPa 5:56  
VENTING 5:56  
AIR DRYING 8:03  
CYCLE COMPLETE 8:05

-----  
Digital Signature #  
AF32807881CFEA2E  
-----