

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 003282  
10:14 24/03/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:29  
PRESSURIZING 1:33  
STERILIZING 2:30  
136.0 C 220kPa 2:30  
136.4 C 224kPa 3:00  
136.2 C 221kPa 3:30  
136.4 C 224kPa 4:00  
136.2 C 221kPa 4:30  
136.5 C 224kPa 5:00  
136.3 C 223kPa 5:30  
136.4 C 221kPa 6:00  
136.4 C 221kPa 6:00  
VENTING 6:00  
AIR DRYING 6:56  
CYCLE COMPLETE 9:08

-----  
Digital Signature #  
C8FD4970C974BEB9  
-----